



SITIIAS
Worldwide Access

防爆合格证

证号：GYB23.1529X

制 造 商 常州市武翔仪表有限公司

(地址：江苏省常州市经济开发区横林镇崔桥街崔蓉路 76 号)

产 品 名 称 隔爆型热电阻

型 号 规 格 aWZbKc-d4Be

防 爆 标 志 Ex db II C T1...T6 Gb, Ex tb III C T80°C Db

产 品 标 准 GB/T 26786-2011

1.184.111~114,

图 样 编 号

WXGBT20190001~20190031,

WXGZC20190001/20190002/20190005/20190007/20190008/20190011/
20190013/20190014/20190017/20190019/20190020/20190023

经图样及技术文件的审查和样品检验，确认上述产品符合下列标准：

GB/T 3836.1-2021, GB/T 3836.2-2021, GB/T 3836.31-2021

特颁发此证。

本证书有效期：2023 年 01 月 09 日 至 2028 年 01 月 08 日

备 注

1. 安全使用注意事项见使用说明书。
2. 证书编号后缀“X”表明产品具有安全使用特殊条件：产品使用环境温度范围：-20°C~+60°C；涉及产品隔爆接合面的维修须联系产品制造厂商。
3. 型号规格说明：a代表热电阻分度号，可为Pt系列；b代表热电阻材料，可为P；c代表感温元件，可为空白或2；d代表连接形式：可为1-无固定、2-螺纹安装和4-法兰安装；e代表温套管结构形式：可为空白（不选用分体套管时此处不标注）或温套管对应代码（A, B, C, D, E, F, G, H, J, K1, K2, K3, L1, L2, L3, M1, M2, M3, N, P, Q, R, S, T, U, HT1, HT2, HT3, HT4, HT5, HT6）。
4. 外壳防护等级 IP65。



批 准

上海仪器仪表自控系统检验测试所有限公司
国家级仪器仪表防爆安全监督检验站


颁发日期二〇二三年一月九日

本证书仅对与认可文件和样品一致的产品有效。

地址：上海市漕宝路103号
邮编：200233

网址：www.nepsi.org.cn
Email: info@nepsi.org.cn

电话：+86 21 64368180
传真：+86 21 64844580



SITIAs
Worldwide Access

EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert No. GYB23.1529X

Manufacturer Changzhou Wuxiang Instrument Co., Ltd.
(Address: №.76 Cuirong Rd, Cuiqiao Street, Henglin Town, Changzhou City, Jiangsu, China)

Product Flameproof Thermal Resistance

Model aWZbKc-d4Be

Ex marking Ex db II C T1...T6 Gb, Ex tb III C T80°C Db

Product standard GB/T 26786-2011
1.184.111~114,
WXGBT20190001~20190031,

Drawing number WXGZC20190001/20190002/20190005/20190007/20190008/20190011/
20190013/20190014/2090017/20190019/20190020/20190023

The product was found to comply with the following standard(s):
GB/T 3836.1-2021,GB/T 3836.2-2021,GB/T 3836.31-2021

Valid until: 2028.01.08

Remarks

- 1.Conditions for safe use are specified in the instructions.
- 2.Symbol "X" placed after the certification number denotes specific conditions of use: Ambient temperature: -20°C~+60°C; Contact the original manufacturer when repair work relates to the flamepath.
- 3.Model designation: **a** indicates index № of thermoresistance, which can be Pt series; **b** indicates material, which can be P; **c** indicates sensor, which can be blank or 2; **d** indicates conection, which can be 1, 2 or 4; **e** indicates protection tube, (A,B,C,D,E,F,G,H,J,K1,K2,K3,L1,L2,L3,M1,M2,M3,N,P,Q,R,S,T,U,HT1,HT2,HT3, HT4,HT5,HT6) .
- 4.IP code: IP65.



Approval

Shanghai Inspection and Testing Institute of
Instruments and Automation Systems Co., Ltd.
National Supervision and Inspection Center for
Explosion Protection and Safety of Instrumentation
Date of issue 2023.01.09

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: info@nepsi.org.cn

Tel: +86 21 64368180
Fax: +86 21 64844580